



CERTIFICATE

CERTIFICATE # IBC 9251310/9265

This is to certify stainless steel medical devices manufactured by

SURGIFORCE INSTRUMENTS

Nasir Road Islamia Park 51310, Sialkot-Pakistan

Has been assessed and found in compliance on requirements for
Surgical and Dental Hand Instruments – Determination of resistance against autoclaving,
corrosion and thermal exposure includes testing for autoclavability, boiling water test, copper
sulfate test and thermal test parameters of standard

ISO 13402:2000

This certificate is valid for the following scope

SURGICAL INSTRUMENTS

DETERMINATION OF RESISTANCE AGAINST AUTOCLAVING, CORROSION AND
THERMAL EXPOSURE FOR

**SURGICAL INSTRUMENTS, DENTAL INSTRUMENTS,
ORTHOPEDIC INSTRUMENTS, MANICURE
INSTRUMENTS, STERILIZATION TRAYS AND BASKETS**

This assessment exercise was carried with all due care, though regular conformance to all the
regulatory requirement is the sole responsibility of manufacturer.

Scheme Manager:

Issue Date: 05-11-2024

Expiry Date: 04-11-2025

Remains valid subject to satisfactory surveillance audit.

To verify the validity of Certificate, please visit www.8ibc.org

Certification is subject to IBC terms and conditions accessible through official web.
Validity may be confirmed via website: www.8ibc.org or email: info@8ibc.org
This certificate remains the property of IBC, to whom it must be returned upon request.

